



校准证书

CALIBRATION CERTIFICATE

证书编号 088201204482
Certificate No.

第 1 页, 共 7 页
Page of

委托方 广州市铁泰电子科技有限公司
Client

委托方地址 广州市白云区嘉禾联边村上致富路F栋4楼
Add. of Client

计量器具名称 3 PHASE DIGITAL PHASE METER
Description

型号规格 4400
Model/Type

制造厂 ETCR
Manufacturer

出厂编号 XWB1111133
Serial No.

设备编号
Equipment No.

接收日期 2012年 2月 3日
Date of Receipt Y M D

结论 见校准结果页
Conclusion Shown in the results of calibration

校准日期 2012年 2月 13日
Date of Calibration Y M D

批准人
Approved Signatory

检验 曹硕
Inspected by

校准 石劲毅
Calibrated by





说 明

证书编号: D88201204482
Certificate No.

DIRECTIONS

第 2 页, 共 7 页
Page of

1. 本中心是国家质量监督检验检疫总局在华南地区设立的国家法定计量检定机构, 计量授权证书号是: (国) 法计(2007) 01043号、(国) 法计(2007) 01032号, 本中心是中国合格评定国家认可委员会 (CNAS) 认可实验室, 认可证书号为: CNAS L0730。

This laboratory is the National Legal Metrological Verification Institution in southern China set up by the General Administration of Quality Supervision, Inspection and Quarantine of the People's Republic of China (AQSIQ) under authorization certificates No.(2007)01043 & (2007)01032. This laboratory is accredited by China National Accreditation Service for Conformity Assessment under Laboratory Accreditation Certification No. CNAS L0730.

2. 本中心所出具的数据均可溯源至国家计量基准和国际单位制 (SI)。

All data issued by this laboratory are traceable to national primary standards and International System of Units (SI).

3. 本次校准的技术依据:

Reference documents for the calibration:

JJG 780-1992 交流数字功率表检定规程 V.R. of AC Digital Powermeter

4. 本次校准所使用的主要计量标准器具:

Major standards of measurement used in the calibration:

设备名称/型号 Name of Equipment /Model	编号 Serial No.	证书号/有效期 Certificate No. /Due Date	计量特性 Metrological Characteristic
三相标准功率电能表 Three-phase Standard Watt-hour Meter /PES200.3	20692	DLdr2011-0321 /2012-04-10	0.02 级 Grade 0.02
功率标准器 Electrical Power Standard Testing Equip /6100A	852147534~85214753 6	DLdr2011-0893 /2012-09-19	0.02 级 Grade 0.02

5. 校准地点、环境条件:

Place and environmental conditions of the calibration:

地点 本中心电磁实验 湿度 (23±2) °C 相对湿度 (50±5) %
Place 国家电测磁学实验 Lab Temperature RH

6. 被校准仪器限制使用条件:

Limiting condition of the instrument calibrated:

注: 1. 本证书校准结果只与受校准仪器有关。

2. 未经本中心书面批准, 不得部分复制此证书。

Note 1. The results relate only to the items calibrated.

2. This certificate shall not be reproduced except in full, without the written approval of our laboratory.



校准结果

RESULTS OF CALIBRATION

证书编号: D00201204482
Certificate No.

原始记录号: 120124484
Record No.

第 3 页, 共 7 页
Page of 7

一、外观: 符合要求(Apparent inspection: Pass)

二、交流电压: 见表1 (ACV: Is table 1, 50 Hz)

表 1 (Table 1)

回路 Channel	量程 Range	示值 Indication Value (V)	实际值 Reference Value (V)	误差 Error (V)	允许误差 MPE (V)	结论 Conclusion (Pass/Fail)
U1	600V	50.1	50.000	+ 0.1	±1.1	Pass
		99	100.00	- 1	±4	Pass
		199	200.00	- 1	±6	Pass
		299	300.00	- 1	±7	Pass
		399	400.00	- 1	±9	Pass
		500	500.00	0	±11	Pass
U2	600V	50.0	50.000	0.0	±1.1	Pass
		100	100.00	0	±5	Pass
		200	200.00	0	±6	Pass
		300	300.00	0	±8	Pass
		400	400.00	0	±9	Pass
		500	500.00	0	±11	Pass
U3	600V	50.0	50.000	0.0	±1.1	Pass
		99	100.00	- 1	±4	Pass
		199	200.00	- 1	±6	Pass
		299	300.00	- 1	±7	Pass
		399	400.00	- 1	±9	Pass
		499	500.00	- 1	±10	Pass



校准结果

RESULTS OF CALIBRATION

证书编号: D68201204482
Certificate No.

原始记录号: 120124484
Record No.

第 4 页, 共 7 页
Page of 7

二、交流电压: 见表2 (ACA: In table 2)/50 Hz)

表 2 (Table 2)

回路 Channel	量程 Range	示值 Indication Value (A)	实际值 Reference Value (A)	误差 Error (A)	允许误差 MPE (A)	结论 Conclusion (Pass/Fail)
11	20A	1.98	2.000	- 0.02	±0.06	Pass
		5.00	5.000	0.00	±0.11	Pass
		10.02	10.000	+ 0.02	±0.18	Pass
		15.07	15.000	+ 0.07	±0.26	Pass
		19.11	19.000	+ 0.11	±0.32	Pass
12	20A	1.98	2.000	- 0.02	±0.06	Pass
		4.99	5.000	- 0.01	±0.10	Pass
		10.02	10.000	+ 0.02	±0.18	Pass
		15.06	15.000	+ 0.06	±0.26	Pass
		19.11	19.000	+ 0.11	±0.32	Pass
13	20A	1.99	2.000	- 0.01	±0.06	Pass
		5.00	5.000	0.00	±0.11	Pass
		10.03	10.000	+ 0.03	±0.18	Pass
		15.07	15.000	+ 0.07	±0.26	Pass
		19.11	19.000	+ 0.11	±0.32	Pass





校准结果

RESULTS OF CALIBRATION

证书编号: D68201204482
Certificate No.

原始记录号: 120124484
Record No.

第 5 页, 共 7 页
Page of 7

四、单相交流功率: 见表3 (Single-phase AC Power: In table 3/50 Hz)

表 3 (Table 3)

回路 Channel	量程 Range	功率因数 PF	示值 Indication Value (kW)	实际值 Reference Value (kW)	误差 Error (kW)	允许误差 MPE (kW)	结论 Conclusion (Pass/Fail)
P1	600V/20A	1.0	0.3960	0.40000	-0.0010	±0.0122	Pass
		1.0	0.9950	1.00000	-0.0013	±0.0302	Pass
		1.0	1.9960	2.00000	-0.0010	±0.0602	Pass
		1.0	2.9989	3.00000	-0.0003	±0.0903	Pass
		1.0	3.8049	3.80000	+0.0012	±0.1144	Pass
		0.5L	1.9015	1.90000	+0.0004	±0.0573	Pass
		0.5C	1.9015	1.90000	+0.0004	±0.0573	Pass
P2	600V/20A	1.0	0.3960	0.40000	-0.0010	±0.0122	Pass
		1.0	0.9980	1.00000	-0.0005	±0.0302	Pass
		1.0	2.0040	2.00000	+0.0010	±0.0604	Pass
		1.0	3.0120	3.00000	+0.0010	±0.0907	Pass
		1.0	3.8220	3.80000	+0.0055	±0.1150	Pass
		0.5L	1.9110	1.90000	+0.0028	±0.0576	Pass
		0.5C	1.9110	1.90000	+0.0028	±0.0576	Pass
P3	600V/20A	1.0	0.3980	0.40000	-0.0005	±0.0122	Pass
		1.0	0.9950	1.00000	-0.0013	±0.0302	Pass
		1.0	1.9980	2.00000	-0.0005	±0.0602	Pass
		1.0	3.0009	3.00000	+0.0002	±0.0903	Pass
		1.0	3.8049	3.80000	+0.0012	±0.1144	Pass
		0.5L	1.9025	1.90000	+0.0006	±0.0574	Pass
		0.5C	1.9025	1.90000	+0.0006	±0.0574	Pass



校准结果 RESULTS OF CALIBRATION

证书编号 088201204482
Certificate No.

原始记录号 120124484
Record No.

第 6 页, 共 7 页
Page of 7

五、相位: 见表4 (Phase: In table 4) $f=50 \text{ Hz}$

表 4 (Table 4)

回路 Channel	示 值 Indication Value (°)	实际值 Reference Value (°)	误 差 Error (°)	允许误差 MPE (%)	结论 Conclusion (Pass/Fail)
U111	0.1	0.00	+ 0.1	±1.0	Pass
	29.8	30.00	- 0.2	±1.0	Pass
	59.8	60.00	- 0.2	±1.0	Pass
	300.0	300.00	0.0	±1.0	Pass
	330.0	330.00	0.0	±1.0	Pass
U112	0.6	0.00	+ 0.6	±1.0	Pass
	30.3	30.00	+ 0.3	±1.0	Pass
	60.3	60.00	+ 0.3	±1.0	Pass
	300.5	300.00	+ 0.5	±1.0	Pass
	330.5	330.00	+ 0.5	±1.0	Pass
U113	29.6	30.00	- 0.4	±1.0	Pass
	59.6	60.00	- 0.4	±1.0	Pass
	299.8	300.00	- 0.2	±1.0	Pass
	329.8	330.00	- 0.2	±1.0	Pass
	359.8	360.00	- 0.2	±1.0	Pass

六、频率: 见表5 (FREQ: In table 5)

表 5 (Table 5)

示 值 Indication Value (Hz)	实际值 Reference Value (Hz)	误 差 Error (Hz)	允许误差 MPE (Hz)	结论 Conclusion (Pass/Fail)
49.96	50.000	- 0.04	±1.03	Pass
59.95	60.000	- 0.05	±1.23	Pass



校准结果

RESULTS OF CALIBRATION

证书编号 088201204482
Certificate No.

原始记录号 120124484
Record No.

第 7 页, 共 7 页
Page of 7

七、功率因数: 见表6 (PF: In table 6, $f=50$ Hz)

表 6 (Table 6)

回路 Channel		示值 Indication Value	实际值 Reference Value	误差 Error	允许误差 MPE	结论 Conclusion
		(cos ϕ)	(cos ϕ)	(cos ϕ)	(cos ϕ)	(Pass/Fail)
PF1	L	0.500	0.5000	0.000	± 0.030	Pass
		0.866	0.8660	0.000	± 0.030	Pass
		1.000	1.0000	0.000	± 0.030	Pass
		0.866	0.8660	0.000	± 0.030	Pass
PF2	L	0.500	0.5000	0.000	± 0.030	Pass
		0.866	0.8660	0.000	± 0.030	Pass
		1.000	1.0000	0.000	± 0.030	Pass
		0.866	0.8660	0.000	± 0.030	Pass
PF3	L	0.500	0.5000	0.000	± 0.030	Pass
		0.866	0.8660	0.000	± 0.030	Pass
		1.000	1.0000	0.000	± 0.030	Pass
		0.866	0.8660	0.000	± 0.030	Pass
PF4	C	0.500	0.5000	0.000	± 0.030	Pass
		0.866	0.8660	0.000	± 0.030	Pass
		1.000	1.0000	0.000	± 0.030	Pass
		0.866	0.8660	0.000	± 0.030	Pass

说明(Notes):

1. 测量结果的扩展不确定度(The Expanded Uncertainty of Measurement):

交流电流(ACA) $U_{95}\% = 0.04\%$ 交流电压(ACV) $U_{95}\% = 0.04\%$

交流功率(AC Power) $U_{95}\% = 0.04\%$ 相位(Phase) $U_{95}\% = 0.3^\circ$

频率(FREQ) $U_{95}\% = 0.010\%$ 功率因数(PF) $U_{95}\% = 0.10\%$

(包含因子 $k=2$ 依据 JJF1059-1999 测量结果不确定度评定与表示)

(Coverage factor $k=2$, According to JJF1059-1999 Evaluation and Expression of Uncertainty in Measurement)

2. 建议校准周期不超过1年。

The period of calibration advised within one year.

