



校准证书

CALIBRATION CERTIFICATE

证书编号 DBB20110606
Certificate No.

第 1 页, 共 3 页
Page of

委托方 广州市依泰电子科技有限公司
Client

委托方地址 广州市白云区嘉禾彭上致富路4号F栋3楼
Add. of Client

计量器具名称 CLAMP LEAKER
Description


型号规格 ETCR6600
Model/Type


制造厂 广州市依泰电子科技有限公司
Manufacturer


出厂编号 LD11060602
Serial No.

接收日期 2011 年 6 月 6 日
Date of Receipt Y M D

校准日期 2011 年 6 月 6 日
Date of Calibration Y M D

批准人 Approved Signatory 

核 验 Inspected by 

校 准 Calibrated by 曹欣 





说 明

证书编号 08820110606
Certificate No.

DIRECTIONS

第 2 页, 共 3 页
Page of

1. 本中心是国家质量监督检验检疫总局在华南地区设立的国家法定计量检定机构, 计量授权证书号是:
(国) 法计 (2007) 01043号、(国) 法计 (2007) 01032号。

This laboratory is the National Legal Metrological Verification Institution in southern China set up by the General Administration of Quality Supervision, Inspection and Quarantine of the People's Republic of China (AQSIQ) under authorization certificates No.(2007)01043 & (2007)01032.

2. 本中心所出具的数据均可溯源至保存在中国计量科学研究院的国家计量基准和国际单位制(SI)。中国计量科学研究院于1999年代表中国签署了“国家计量基准及国家计量研究院出具的校准和测量证书相互承认协议”。All data issued by this laboratory are traceable to national primary standards maintained in National Institute of Metrology (NIM) and International System of Units (SI). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.

3. 本次校准的技术依据:

Reference documents for the calibration:

JJF1075-2001 钳形电流表校准规范 C. S. of Clamp Ammeters

4. 本次校准所使用的主要计量标准器具:

Major standards of measurement used in the calibration:

设备名称/型号 Name of Equipment /Model	编号 Serial No.	证书号/有效期 Certificate No. /Due Date	计量特性 Metrological Characteristic
多功能标准源 Multi-function Standard Sources /5520A	8080014	08820110606 /2011-06-06	DCV: $\pm 0.0011\%$ DCA: $\pm 0.01\%$ ACV: $\pm 0.015\%$ ACA: $\pm 0.04\%$ OHM: $\pm 0.0028\%$ FREQ: $\pm (2.5 \times 10^{-6})$

5. 校准地点、环境条件:

Place and environmental conditions of the calibration:

地点 本中心电磁实验室(Electrics Lab) 温度 $(23 \pm 2) ^\circ\text{C}$ 相对湿度 $(55 \pm 5) \% \text{RH}$

6. 被校准仪器限制使用条件:

Limiting condition of the instrument calibrated:

注: 1. 本证书校准结果只与受校准仪器有关。
2. 未经本中心书面批准, 不得部分复制此证书。
Note 1. The results relate only to the items calibrated.
2. This certificate shall not be reproduced except in full, without the written approval of our laboratory.

2. This certificate shall not be reproduced except in full, without the written approval of our laboratory.



校准结果 RESULTS OF CALIBRATION

证书编号 DBB20110606
Certificate No.

原始记录号 120091013
Record No.

第 3 页, 共 3 页
Page of

一、外观(Appearance): 符合要求(Pass)

二、交流电流: 见表1 (ACA: In table 1, f=50 Hz)

表 1 (Table 1)

量 程 Range	示 值 Indication Value (mA)	实 际 值 Reference Value (mA)	误 差 Error (mA)	允 许 误 差 MPE (mA)	结 论 conclusion (Pass/Fail)
30 mA	3.01	3.000	+ 0.01	±0.09	Pass
	15.1	15.00	+ 0.1	±0.7	Pass
	30.3	30.00	+ 0.3	±1.0	Pass
300 mA	50.0	50.00	0.0	±1.3	Pass
	99.5	100.00	- 0.5	±2.0	Pass
	150.5	150.00	+ 0.5	±2.8	Pass
	200.7	200.00	+ 0.7	±3.5	Pass
	300.3	300.00	+ 0.3	±5.0	Pass
30 A	(A)	(A)	(A)	(A)	
	3.00	3.000	0.00	±0.09	Pass
	14.9	15.00	- 0.1	±0.7	Pass
300 A	29.8	30.00	- 0.2	±0.9	Pass
	50.1	50.00	+ 0.1	±1.3	Pass
	152.6	150.00	+ 2.6	±3.6	Pass
	293.1	300.00	- 6.9	±12.2	Pass

说明(Notes):

1. 测量结果的相对扩展不确定度(The Relative Expanded Uncertainty of Measurement):

交流电流(ACA) $U_{rel}=0.13\%$

(包含因子 $k=2$ 依据 JJF1059-1999 测量结果不确定度与表示)

(Coverage factor $k=2$. According to JJF1059-1999 Evaluation and Expression of Uncertainty in Measurement)

2. 建议校准周期不超过1年

The period of calibration advised within one year.