

Certificate of Conformity

Registration No.: T1850927 01

Applicant : ETCR Electronic Technology Company
Address : F-3F, No.4 Pengshang Zhifu Road, Jiahe, Baiyun District,
Guangzhou, Guangdong, China.
Product : Split Type High Accuracy Leakage current Sensor
Model No. : ETCR010K, ETCR010KD, ETCR025K, ETCR025KD,
ETCR040K, ETCR040KD, ETCR040KG, ETCR085K

Trademark : 

The submitted products have been tested by us with the listed standards and found in compliance with the following European Directives:

The EMC Directive 2004/108/EC

EN 61326-1:2013

EN 61326-2-2:2013

The tests were performed in normal operation mode, the test results apply only to the particular sample tested and to the specific tests carried out. This certificate applies specifically to the sample investigated in our test reference number only.

The CE markings as shown below can be affixed on the product after preparation of necessary technical documentation.

Other relevant Directives have to be observed.



Certified by:

August 06, 2015



Shenzhen Alpha Product Testing Co., Ltd.

Building B, East Area of Nanchang Second Industrial Zone,

Gushu 2nd Road, Bao'an District, Shenzhen 518126, P. R. China

Website: <http://www.a-lab.cn> Email: service@a-lab.cn

Certificate of Conformity

Registration No.: T1850927 02

Applicant name : ETCR Electronic Technology Company

Address : F-3F, No.4 Pengshang Zhifu Road, Jiahe, Baiyun District, Guangzhou, Guangdong, China.

Product : Split Type High Accuracy Leakage current Sensor

Model No. : ETCR010K, ETCR010KD, ETCR025K, ETCR025KD, ETCR040K, ETCR040KD, ETCR040KG, ETCR085K

Trademark : 

Ratings :

Max. measuring current:	60A
Max. Line voltage:	600V

The submitted products have been tested by us with the listed standards and found in compliance with the following European Directives:

The Low Voltage Directive 2006/95/EC

EN 61010-1: 2010
EN 61010-2-032: 2012

The tests were performed in normal operation mode, the test results apply only to the particular sample tested and to the specific tests carried out. This certificate applies specifically to the sample investigated in our test reference number only.

The CE markings as shown below can be affixed on the product after preparation of necessary technical documentation.

Other relevant Directives have to be observed.



Certified by: _____

July 30, 2015



Shenzhen Alpha Product Testing Co., Ltd.

Building B, East Area of Nanchang Second Industrial Zone,
Gushu 2nd Road, Bao'an District, Shenzhen 518126, P. R. China
Website: <http://www.a-lab.cn> Email: service@a-lab.cn